

## IMPROVED PCB CONGENERS SPECIFIC ANALYSIS BY HRGC-HRMS

Pond P., Hope D., Ma E.

Pacific Rim Laboratories Inc., BC Canada, V3S 8P8

### Abstract

A study was proposed to simultaneously monitor all 209 PCB congeners in various matrices. A state of the art DFS HRGC-HRMS coupled with a newly developed SGE column were employed in achieving lower detection limits and better chromatographic separation. The results indicated detection limits for tissue samples were 1~10 ng/kg (ppt), 10 pg/L for water samples. A total of one hundred & eighty nine peaks were resolved, consisting of 172 individual congeners and 17 coelutions.

### Introduction

Environmental samples, especially tissue, do not give results that resemble Aroclors, therefore they are difficult to quantify based on traditional methods<sup>1</sup>. Hence, congener speciation has to be achieved to fulfill various analysis requests. The study was carried out using U.S. EPA Method 1668B<sup>2</sup> as our reference.

### Materials and Methods

HRGC-HRMS congener specific method uses 27 internal standards (<sup>13</sup>C<sub>12</sub>-labelled congeners) which are: all twelve dioxin-like PCBs (dlPCBs), first and last elute congener in each level of chlorination (LOC); three <sup>13</sup>C<sub>12</sub>-labelled clean-up standards and five <sup>13</sup>C<sub>12</sub>-labelled recovery standards. A six point (0.2~2000 ng/mL) calibration was created for the quantification of dlPCBs and LOC PCBs. A single point calibration was used for all remaining PCB congeners.

A newly developed SGE GC column was adopted for better congener chromatographic separation in our study. SGE HT8 PCB, 8% phenyl (equiv.) polycarborane-siloxane, 60m, 0.25mm i.d., 0.15um film thickness.

A Thermo DFS HRMS with dual Thermo Trace GCs with split/splitless injection was used with the following conditions: Initial temp 90 °C for 1 min, ramp 20 °C/min to 190 °C, then ramp 1.5 °C/min to 250 °C and ramp 4.0 °C/min to 300 °C for 4.2 mins.

### Results and Discussion

Fig. 1 demonstrates from mono- to deca-CBs, 189 out of 209 PCB congeners or congener pairs were resolved. Table 1 lists the retention time of the congener peaks in this study. All dioxin-like PCBs were resolved from other congeners in the given level of chlorination. PCB105 has shoulder separation from PCB127, however the latter is a rarely occurring congener. PCB110 gives a fragment ion (M-Cl) that resolves from PCB081, however high concentrations of PCB110 in environmental samples may mask the low concentrations usually found for PCB081.

Fig.2 shows excellent sensitivity of 0.2 pg/μL PCB Std solution on DFS HRMS. Fig.3 (a)(b)(c) indicates better chromatographic resolution with SGE HT8 PCB column. The combination of the SGE column resolution and DFS HRMS sensitivity allows for a more dilute (200 μL) final sample volume. This leads to less environmental matrix effects, thus increasing the system stability and ultimately sample throughput. Fig.4 (a)(b) demonstrates the results from a fish tissue sample. During the course of six weeks, and over 250 injections of fish extracts, the column maintained its chromatographic resolution. Fig.5 (a)(b)(c) shows the results from serum samples.

### Acknowledgements

Pacific Rim Laboratories would like to thank Wellington Laboratories with gratitude for their long-term support. And in appreciation of SGE in collaboration of column performance testing and study.

### References

1. Bernhard T., Petron S., *PCB Congeners in Ecological Risk Assessment*, 2001
2. U.S. EPA, *Method 1668B* 2008.

Table 1 The retention time of the congeners peaks in this study

Congener Number	9 Pt Std	CS209	Congener Number	9 Pt Std	CS209	Congener Number	9 Pt Std	CS209	Congener Number	9 Pt Std	CS209	Congener Number	9 Pt Std	CS209
PCB-001	10.38	10.29	PCB-054	25.37	25.37	PCB-104	33.92	33.90	PCB-155	43.35	43.37	PCB-188	55.30	55.31
PCB-002	13.02	12.94	PCB-050	27.37	27.35	PCB-096	37.12	37.15	PCB-150	46.70	46.70	PCB-184	56.25	56.27
PCB-003	13.30	13.23	PCB-053	28.95	28.98	PCB-103	38.48	38.50	PCB-152	47.77	47.77	PCB-179	57.48	57.50
			PCB-051	29.71	29.73	PCB-100	39.55	39.55	PCB-145	50.01	49.99	PCB-176	58.31	58.31
			PCB-045	30.56	30.56	PCB-094	40.55	40.59	PCB-136	50.05	50.06	PCB-186	58.91	58.93
PCB-010	14.18	14.04	PCB-046	32.09	32.09	PCB-093	41.75		PCB-148	50.58	50.58	PCB-178	59.98	60.00
PCB-004	14.29	14.22	PCB-069	32.90	33.01	PCB-102	41.89	41.81	PCB-154	51.88	51.89	PCB-175	60.39	60.41
PCB-009	16.35	16.31	PCB-062	33.01		PCB-098	41.96		PCB-151	52.99	52.99	PCB-182	60.54	60.56
PCB-007	16.60	16.49	PCB-073	33.27	33.31	PCB-095	42.11		PCB-135	53.51	53.52	PCB-187	60.56	
PCB-006	17.78	17.76	PCB-043	33.45	33.68	PCB-088	42.55	42.59	PCB-144	53.73	53.87	PCB-183	60.94	60.95
PCB-008	18.31	18.30	PCB-049	33.68		PCB-091	43.15	43.15	PCB-147	53.80	53.81	PCB-185	61.38	61.38
PCB-005	18.39		PCB-065	33.95	33.94	PCB-121	43.69	43.73	PCB-149	54.56	54.59	PCB-174	61.66	61.66
PCB-014	19.95	19.88	PCB-075	34.23		PCB-084	45.76	45.80	PCB-139	54.59		PCB-181	61.74	61.74
PCB-011	22.45	22.47	PCB-062	34.39	34.27	PCB-092	45.98	45.99	PCB-140	55.08	55.10	PCB-177	61.90	61.89
PCB-013	23.17	23.16	PCB-048	34.35		PCB-089	46.18	46.22	PCB-143	55.41	55.43	PCB-171	62.19	62.20
PCB-012	23.19		PCB-047	34.40	34.38	PCB-090	46.84	46.88	PCB-134	55.61	55.63	PCB-173	62.40	62.40
PCB-015	24.02	23.96	PCB-044	36.33	36.35	PCB-101	47.17	47.19	PCB-142	56.03	56.02	PCB-172	62.99	62.99
			PCB-059	36.56	36.59	PCB-113	47.65	47.68	PCB-131	56.25	56.27	PCB-192	63.13	63.14
			PCB-042	36.98	36.99	PCB-099	48.10	48.12	PCB-133	56.70	56.70	PCB-180	63.32	63.35
PCB-019	19.56	19.50	PCB-064	37.85	37.86	PCB-112	49.10	49.12	PCB-165	57.25	57.25	PCB-193	63.35	63.45
PCB-030	20.41	20.38	PCB-072	38.27		PCB-119	49.34		PCB-132	57.45	57.50	PCB-191	63.55	63.57
PCB-018	22.27	22.25	PCB-071	38.28	38.18	PCB-083	49.59	49.72	PCB-146	57.50	57.50	PCB-170	64.24	64.24
PCB-017	22.67	22.61	PCB-041	38.37	38.33	PCB-109	49.73	49.72	PCB-161	57.78	57.78	PCB-190	64.35	64.38
PCB-024	23.24	23.21	PCB-068	39.08	39.11	PCB-086	50.31		PCB-153	58.29	58.32	PCB-189	65.67	65.68
PCB-027	23.75	23.73	PCB-040	39.81	39.83	PCB-125	50.43	50.48	PCB-168	58.39	58.39			
PCB-032	24.66	24.62	PCB-057	40.37	40.40	PCB-117	50.48		PCB-141	59.06	59.08			
PCB-016	24.90	24.88	PCB-067	41.32	41.35	PCB-097	50.63	50.91	PCB-137	59.51	59.52	PCB-202	61.42	61.45
PCB-023	25.95	25.95	PCB-063	41.69	41.73	PCB-116	50.87		PCB-130	59.70	59.70	PCB-200	61.85	61.87
PCB-034	26.41	26.39	PCB-058	41.92	41.93	PCB-115	51.11	51.16	PCB-163	60.06	60.08	PCB-204	61.97	61.97
PCB-029	26.61	26.59	PCB-061	42.18	42.20	PCB-087	51.19		PCB-164	60.07		PCB-197	62.26	62.27
PCB-026	27.56	27.54	PCB-074	42.73	42.78	PCB-111	51.78	51.89	PCB-138	60.19	60.21	PCB-201	62.79	62.81
PCB-025	27.99	27.98	PCB-070	43.54	43.59	PCB-095	51.91	51.99	PCB-160	60.21	60.31	PCB-198	64.05	64.08
PCB-031	28.50	28.49	PCB-055	43.78	43.78	PCB-110	52.60	52.65	PCB-158	60.37	60.39	PCB-199	64.10	64.13
PCB-028	29.03	29.03	PCB-066	44.36	44.39	PCB-120	52.68	52.73	PCB-129	60.53	60.53	PCB-196	64.37	64.39
PCB-021	29.98	29.99	PCB-080	44.37	44.49	PCB-082	54.02	54.05	PCB-166	60.79	60.79	PCB-203	64.41	64.43
PCB-020	30.55	30.56	PCB-076	45.58	45.61	PCB-124	55.49	55.51	PCB-159	61.47	61.49	PCB-195	65.37	65.36
PCB-033	30.59		PCB-060	46.94	47.00	PCB-108	55.81	55.86	PCB-128	61.59	61.61	PCB-194	66.60	66.60
PCB-022	31.49	31.51	PCB-056	47.22	47.26	PCB-107	55.84		PCB-162	61.74	61.76	PCB-205	66.96	66.95
PCB-036	32.70	32.73	PCB-079	49.75	49.84	PCB-123	56.14	56.15	PCB-167	62.09	62.09			
PCB-039	33.78	33.85	PCB-078	51.27	51.36	PCB-106	56.28	56.30	PCB-156	62.98	62.97			
PCB-038	35.41	35.43	PCB-081	52.51	52.59	PCB-118	56.59	56.61	PCB-157	63.17	63.17	PCB-208	64.69	64.72
PCB-035	37.36	37.43	PCB-077	54.16	54.20	PCB-114	57.27	57.29	PCB-169	64.43	64.46	PCB-207	65.04	65.05
PCB-037	38.64	38.72				PCB-122	57.87	57.91				PCB-206	67.59	67.59
						PCB-105	59.08	59.09						
						PCB-127	59.26	59.28						
						PCB-126	61.33	61.36				PCB-209	68.05	68.05




Fig. 1 Mono- to deca-CBs




Fig. 2(a) PCB CS-Lo Std at 0.2 pg/uL




Fig. 2(b) 6 point calibration results at 0.2 ~2000 ng/mL




Fig.3 (a)(b)(c) Chromatographic resolution with SGE HT8 PCB column





Fig.4 (a)(b) The results from fish tissue sample






Fig.5 (a)(b)(c) The results from a serum sample